



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: May 29, 2012 INITIALS: DEC OWNER: WASHINGTON STATE PATROL

DISTRICT: 5 ADDRESS: SR503 MP49

LOCATION: Woodland 83

COUNTY: Cowlitz

MAKE: Rice Lake

TYPE: Trk Scale

ACCURACY CLASS: III L

MODEL: IQ 310

ID. NO: 98438451

CAPACITY: 100000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 20 lbs.

REG. ELEMENT: Electronic

PLATFORM 8x10

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20000	20000	0				
CENTER	20000	20000	0				
SEC 2	20000	20020	+20				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L3184-2 and 200446-0-L3184-4

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.


Signed by:

David Cromer

Date:

June 14, 2012

Print name: David Cromer



State of Washington

County of Thurston

Signed and sworn to (or affirmed before me on) June 14, 2012

Patricia A. Geppert

(Signature of notary public)

Patricia A. Geppert

Printed Name of Notary

My appointment expires Sept. 4, 2015